



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Nissenken Quality Evaluation Center OEKO-TEX® Laboratory***  
***2-16-11 Kuramae Taito-ku, Tokyo 111-0051***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Analysis of Pb in Metal, Nonmetal and Paint, Detection test of 24 certain aromatic amines in textile products and resin products, and Detection test of 8 phthalate esters in textile products and resin products***  
***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President/Operations Manager

*Initial Accreditation Date:*

May 15, 2012

*Issue Date:*

May 7, 2020

*Expiration Date:*

May 31, 2022

*Accreditation No.:*

71517

*Certificate No.:*

L20-265

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjlabs.com](http://www.pjlabs.com)*



# Certificate of Accreditation: Supplement

**Nissenken Quality Evaluation Center OEKO-TEX® Laboratory**

2-16-11 Kuramae Taito-ku, Tokyo 111-0051

Contact Name: Toshiaki Yamazaki Phone: 03-5809-2810

*Accreditation is granted to the facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical <sup>F</sup>	Metal, Nonmetal	Pb	“Procedures for lead-containing Test”	Reference value by Cabinet Order 100 mg/kg
	Paint		On the basis of : 16 CFR Part 1303 CPSC-CH-E1001-08(Metal Products), CPSC-CH-E1002-08(Non-Metal Products), CPSC-CH-E1003-09 (Paint & Coatings) ICP Optical Emission Spectrometry	Reference value by Cabinet Order 90 mg/kg
Textile products Resin products Coloring	24 certain aromatic amines	4-aminobiphenyl	“Procedures for Certain Aromatic Amines Test (JIS L 1940-1)”	Reference value by Cabinet Order 30 mg/kg
		benzidine	On the basis of: Textiles-Methods for determination of certain aromatic amines derived from azo colorants- Part 1: Detection of the use of certain azo colorants accessible with and without extracting the fibres And “Procedures for 4-AAB Test (JIS L 1940-3)”  On the basis of: Textiles-Methods for determination of certain aromatic amines derived from azo colorants-Part 3: Detection of the use of certain azo colorants, which may release 4-aminoazobenzene  GCMS LCMS	
		4-chloro-o-toluidine		
		2-naphthylamine		
		o-aminoazotoluene		
		2-amino-4-nitrotoluene		
		p-chloroaniline		
		2,4-diaminoanisole		
		4,4'-diaminodiphenylmethane		
		3,3'-dichlorobenzidine		
		3,3'-dimethoxybenzidine		
		3,3'-dimethylbenzidine		
		3,3'-dimethyl-4,4'-diaminobiphenylmethane		
		p-cresidine		
		4,4'-methylene-bis-(2-chloroaniline)		
		4,4'-oxydianiline		
		4,4'-thiodianiline		
		o-toluidine		
		2,4-toluylenediamine		
		2,4,5-trimethylaniline		
	o-anisidine			
	2,4-xylydine			
	2,6-xylydine			
	4-aminoazobenzene			



# Certificate of Accreditation: Supplement

**Nissenken Quality Evaluation Center OEKO-TEX® Laboratory**

2-16-11 Kuramae Taito-ku, Tokyo 111-0051

Contact Name: Toshiaki Yamazaki Phone: 03-5809-2810

*Accreditation is granted to the facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical <sup>F</sup>	Textile products Resin products	8 phthalate esters	“Procedures for CPSC Phthalate Test (CPSC-CH-C1001-09.4)”	Reference value by Cabinet Order 1 000 mg/kg
		Dibutyl phthalate (DBP)		
		Di (2-ethylhexyl) phthalate (DEHP)	On the basis of: CPSC-CH-C1001-09.4	
		Benzyl Butyl Phthalate (BBP)		
		Diisononyl phthalate (DINP)	GCMS	
		Diisobutyl Phthalate (DIBP)		
		Di-n-pentylphthalate (DPENP)		
		Di-n-hexyl phthalate (DHEXP)		
		Dicyclohexyl phthalate (DCHP)		

1. The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer F would mean that the laboratory performs this testing at its fixed location.